

**BUNNELL #11 EXTRA SMALL
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

Authorized Representative:
Medica Surgical Innovations Limited
Moorgate Street
Blackburn BB2 4PB
Lancs UK



Single patient use

CAUTION: Federal Law restricts this device
to sale by or on the order of a physician

Printed on Recycled Paper



**BUNNELL #11 EXTRA SMALL
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

Authorized Representative:
Medica Surgical Innovations Limited
Moorgate Street
Blackburn BB2 4PB
Lancs UK



Single patient use

CAUTION: Federal Law restricts this device
to sale by or on the order of a physician

Printed on Recycled Paper



**BUNNELL #11 EXTRA SMALL
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

Authorized Representative:
Medica Surgical Innovations Limited
Moorgate Street
Blackburn BB2 4PB
Lancs UK



Single patient use

CAUTION: Federal Law restricts this device
to sale by or on the order of a physician

Printed on Recycled Paper



**BUNNELL #11 EXTRA SMALL
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

Authorized Representative:
Medica Surgical Innovations Limited
Moorgate Street
Blackburn BB2 4PB
Lancs UK



Single patient use

CAUTION: Federal Law restricts this device
to sale by or on the order of a physician

Printed on Recycled Paper



**BUNNELL #11 EXTRA SMALL
Finger Knuckle Bender Splint**

- >Finger Knuckle Bender Splint to flex proximal interphalangeal joint.
- >Knuckle bender spalk voor vinger voor flexie van het proximale interfalangeale gewricht.
- >Attelle de flexion de l'articulation interphalangienne proximale.
- >Elastischer Quengel zur Beugung des Finger Mittelgelenkes.
- >Stecca pieganocche delle dita per la flessione delle articolazioni interfalangee prossimali.
- >Entablillado doblador de los nudillos de los dedos para doblar la coyuntura interfalangial proximal.

**BUNNELL #11 EXTRA SMALL
Finger Knuckle Bender Splint**

- >Finger Knuckle Bender Splint to flex proximal interphalangeal joint.
- >Knuckle bender spalk voor vinger voor flexie van het proximale interfalangeale gewricht.
- >Attelle de flexion de l'articulation interphalangienne proximale.
- >Elastischer Quengel zur Beugung des Finger Mittelgelenkes.
- >Stecca pieganocche delle dita per la flessione delle articolazioni interfalangee prossimali.
- >Entablillado doblador de los nudillos de los dedos para doblar la coyuntura interfalangial proximal.

**BUNNELL #11 EXTRA SMALL
Finger Knuckle Bender Splint**

- >Finger Knuckle Bender Splint to flex proximal interphalangeal joint.
- >Knuckle bender spalk voor vinger voor flexie van het proximale interfalangeale gewricht.
- >Attelle de flexion de l'articulation interphalangienne proximale.
- >Elastischer Quengel zur Beugung des Finger Mittelgelenkes.
- >Stecca pieganocche delle dita per la flessione delle articolazioni interfalangee prossimali.
- >Entablillado doblador de los nudillos de los dedos para doblar la coyuntura interfalangial proximal.

**BUNNELL #11 EXTRA SMALL
Finger Knuckle Bender Splint**


- >Finger Knuckle Bender Splint to flex proximal interphalangeal joint.
- >Knuckle bender spalk voor vinger voor flexie van het proximale interfalangeale gewricht.
- >Attelle de flexion de l'articulation interphalangienne proximale.
- >Elastischer Quengel zur Beugung des Finger Mittelgelenkes.
- >Stecca pieganocche delle dita per la flessione delle articolazioni interfalangee prossimali.
- >Entablillado doblador de los nudillos de los dedos para doblar la coyuntura interfalangial proximal.

INSTRUCTIONS #11 Finger Knuckle Bender Splint

Applying splint with only one end of the rubber band attached to hook (no dynamic force to flex P.I.P. joint) will facilitate proper positioning of splint on finger. Slide finger into splint, position wider finger band on top of finger and behind D.I.P. joint. Position bar under P.I.P. joint. Complete attaching rubber bands. Evaluate fit, function, and instruct patient on use. The splint is designed to dynamically flex finger while allowing active extension. Rubber bands can be added or subtracted and the wires and bands can be bent to modify fit and function. Remove splint from patient to modify.

Disconnecting rubber bands will make splint removal easier and prepare splint for reapplication.

Clean with mild soap and damp towel. Air dry completely before reapplying. Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritation, or have any questions or concerns.




Bunnell #11 Finger Knuckle Bender to flex proximal interphalangeal joint

INSTRUCTIONS #11 Finger Knuckle Bender Splint

Applying splint with only one end of the rubber band attached to hook (no dynamic force to flex P.I.P. joint) will facilitate proper positioning of splint on finger. Slide finger into splint, position wider finger band on top of finger and behind D.I.P. joint. Position bar under P.I.P. joint. Complete attaching rubber bands. Evaluate fit, function, and instruct patient on use. The splint is designed to dynamically flex finger while allowing active extension. Rubber bands can be added or subtracted and the wires and bands can be bent to modify fit and function. Remove splint from patient to modify.

Disconnecting rubber bands will make splint removal easier and prepare splint for reapplication.

Clean with mild soap and damp towel. Air dry completely before reapplying. Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritation, or have any questions or concerns.




Bunnell #11 Finger Knuckle Bender to flex proximal interphalangeal joint

INSTRUCTIONS #11 Finger Knuckle Bender Splint

Applying splint with only one end of the rubber band attached to hook (no dynamic force to flex P.I.P. joint) will facilitate proper positioning of splint on finger. Slide finger into splint, position wider finger band on top of finger and behind D.I.P. joint. Position bar under P.I.P. joint. Complete attaching rubber bands. Evaluate fit, function, and instruct patient on use. The splint is designed to dynamically flex finger while allowing active extension. Rubber bands can be added or subtracted and the wire and bands can be bent to modify fit and function. Remove splint from patient to modify.

Disconnecting rubber bands will make splint removal easier and prepare splint for reapplication.

Clean with mild soap and damp towel. Air dry completely before reapplying. Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritation, or have any questions or concerns.




Bunnell #11 Finger Knuckle Bender to flex proximal interphalangeal joint

INSTRUCTIONS #11 Finger Knuckle Bender Splint

Applying splint with only one end of the rubber band attached to hook (no dynamic force to flex P.I.P. joint) will facilitate proper positioning of splint on finger. Slide finger into splint, position wider finger band on top of finger and behind D.I.P. joint. Position bar under P.I.P. joint. Complete attaching rubber bands. Evaluate fit, function, and instruct patient on use. The splint is designed to dynamically flex finger while allowing active extension. Rubber bands can be added or subtracted and the wires and bands can be bent to modify fit and function. Remove splint from patient to modify.

Disconnecting rubber bands will make splint removal easier and prepare splint for reapplication.

Clean with mild soap and damp towel. Air dry completely before reapplying. Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritation, or have any questions or concerns.



Bunnell #11 Finger Knuckle Bender to flex proximal interphalangeal joint