

**BUNNELL #6 EXTRA SMALL  
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA  
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

**Authorized Representative:**

Medica Surgical Innovations Limited  
Moorgate Street  
Blackburn BB2 4PB  
Lancs UK

CAUTION: Federal Law restricts this device  
to sale by or on the order of a physician

Printed on Recycled Paper



Single patient use



**BUNNELL #6 EXTRA SMALL  
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA  
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

**Authorized Representative:**

Medica Surgical Innovations Limited  
Moorgate Street  
Blackburn BB2 4PB  
Lancs UK

CAUTION: Federal Law restricts this device  
to sale by or on the order of a physician

Printed on Recycled Paper



Single patient use



**BUNNELL #6 EXTRA SMALL  
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA  
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

**Authorized Representative:**

Medica Surgical Innovations Limited  
Moorgate Street  
Blackburn BB2 4PB  
Lancs UK

CAUTION: Federal Law restricts this device  
to sale by or on the order of a physician

Printed on Recycled Paper



Single patient use



**BUNNELL #6 EXTRA SMALL  
WENIGER**

915 Industrial Way, San Jacinto, CA 92582 USA  
PH: (951) 654-2333 FAX: (951) 654-2331

www.BunnellSplints.com

**Authorized Representative:**

Medica Surgical Innovations Limited  
Moorgate Street  
Blackburn BB2 4PB  
Lancs UK

CAUTION: Federal Law restricts this device  
to sale by or on the order of a physician

Printed on Recycled Paper



Single patient use



**BUNNELL #6 EXTRA SMALL  
Safety Pin Splint**

- >Spring wire Safety Pin Splint to extend a finger.
- >Safety Pin Spalk met veer voor vinger extensie.
- >Attelle d'extension pour un doigt à ressort métallisé et munie d'une épingle de sûreté.
- >Drahtfeder Sicherheitsnadel Splint zur Streckung eines Langfingers.
- >Stecca a spilla di sicurezza con elastico per l'estensione delle dita.
- >Entablillado de resorte de alambre con clavija de seguridad para extender un dedo.

**BUNNELL #6 EXTRA SMALL  
Safety Pin Splint**

- >Spring wire Safety Pin Splint to extend a finger.
- >Safety Pin Spalk met veer voor vinger extensie.
- >Attelle d'extension pour un doigt à ressort métallisé et munie d'une épingle de sûreté.
- >Drahtfeder Sicherheitsnadel Splint zur Streckung eines Langfingers.
- >Stecca a spilla di sicurezza con elastico per l'estensione delle dita.
- >Entablillado de resorte de alambre con clavija de seguridad para extender un dedo.

**BUNNELL #6 EXTRA SMALL  
Safety Pin Splint**

- >Spring wire Safety Pin Splint to extend a finger.
- >Safety Pin Spalk met veer voor vinger extensie.
- >Attelle d'extension pour un doigt à ressort métallisé et munie d'une épingle de sûreté.
- >Drahtfeder Sicherheitsnadel Splint zur Streckung eines Langfingers.
- >Stecca a spilla di sicurezza con elastico per l'estensione delle dita.
- >Entablillado de resorte de alambre con clavija de seguridad para extender un dedo.

**BUNNELL #6 EXTRA SMALL  
Safety Pin Splint**

- >Spring wire Safety Pin Splint to extend a finger.
- >Safety Pin Spalk met veer voor vinger extensie.
- >Attelle d'extension pour un doigt à ressort métallisé et munie d'une épingle de sûreté.
- >Drahtfeder Sicherheitsnadel Splint zur Streckung eines Langfingers.
- >Stecca a spilla di sicurezza con elastico per l'estensione delle dita.
- >Entablillado de resorte de alambre con clavija de seguridad para extender un dedo.


#### INSTRUCTIONS #6 Safety Pin Splint to straighten P.I.P. Joint

Open strap. Place finger into splint with curved finger band on bottom and behind D.I.P. joint. Align strap with P.I.P. joint, position pad on strap (hook and loop compatible), pass strap over top of finger, under lateral wire and back on top of finger. Secure to loop. Evaluate fit, function and instruct patient on use.

Tightening the strap applies a force to straighten finger. Adjust as needed. Lateral wires and finger bands can be carefully reshaped to improve fit and function. Remove from patient to modify.

Clean with mild soap and damp towel. Air dry completely before reapplying.

Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritations or have any questions or concerns.



Bunnell #6 Spring Wire Safety Pin Splint to extend a finger.


#### INSTRUCTIONS #6 Safety Pin Splint to straighten P.I.P. Joint

Open strap. Place finger into splint with curved finger band on bottom and behind D.I.P. joint. Align strap with P.I.P. joint, position pad on strap (hook and loop compatible), pass strap over top of finger, under lateral wire and back on top of finger. Secure to loop. Evaluate fit, function and instruct patient on use.

Tightening the strap applies a force to straighten finger. Adjust as needed. Lateral wires and finger bands can be carefully reshaped to improve fit and function. Remove from patient to modify.

Clean with mild soap and damp towel. Air dry completely before reapplying.

Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritations or have any questions or concerns.



Bunnell #6 Spring Wire Safety Pin Splint to extend a finger.


#### INSTRUCTIONS #6 Safety Pin Splint to straighten P.I.P. Joint

Open strap. Place finger into splint with curved finger band on bottom and behind D.I.P. joint. Align strap with P.I.P. joint, position pad on strap (hook and loop compatible), pass strap over top of finger, under lateral wire and back on top of finger. Secure to loop. Evaluate fit, function and instruct patient on use.

Tightening the strap applies a force to straighten finger. Adjust as needed. Lateral wires and finger bands can be carefully reshaped to improve fit and function. Remove from patient to modify.

Clean with mild soap and damp towel. Air dry completely before reapplying.

Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritations or have any questions or concerns.



Bunnell #6 Spring Wire Safety Pin Splint to extend a finger.


#### INSTRUCTIONS #6 Safety Pin Splint to straighten P.I.P. Joint

Open strap. Place finger into splint with curved finger band on bottom and behind D.I.P. joint. Align strap with P.I.P. joint, position pad on strap (hook and loop compatible), pass strap over top of finger, under lateral wire and back on top of finger. Secure to loop. Evaluate fit, function and instruct patient on use.

Tightening the strap applies a force to straighten finger. Adjust as needed. Lateral wires and finger bands can be carefully reshaped to improve fit and function. Remove from patient to modify.

Clean with mild soap and damp towel. Air dry completely before reapplying.

Immediately contact your health care professional if you are experiencing loss of sensation, swelling, skin irritations or have any questions or concerns.



Bunnell #6 Spring Wire Safety Pin Splint to extend a finger.